

F-QC-001 Rev.00

Certificate Of Analysis

Product Name : Buffer Solution pH 10.0 (Borate)

Catalog Number : L-1202

Batch Number : 1202-00000000

CAS Number : Not applicable to mixtures

Molecular Weight: Not applicable to mixture

Mol. Formula : Not applicable to mixtures

Expire Date : 24 Months



No	Test	Specifications	Results
1	Appearance	Clear colourless liquid	Clear colourless liquid
2	pH at 25 °C	9.95± 0.05	10.03
3	Wt per ml at 20 °C	~ 1.01 g/cm ³	1.01 g/cm ³

LABOTIQ Product specification compared with Buffer standard Solution traceable to NIST



SUDIRO S.Si.
Manager QC