

## Certificate Of Analysis

Product Name : McFarland Standard 0.5

Catalog Number : L-1237

Batch Number : 1237 -00000000

CAS Number : Not Available

Molecular weight : Not Available

Mol.Formula : Not Available

Expire date : 24 Months



No	Test	Specifications	Results
1	Appearance	A colorless to white liquid with a slight bluish tinge, which is created by a fine, white suspension of barium sulfate precipitate	conforms
2	Absorbance (at $\lambda$ 625 nm, 1cm)	0.08 – 0.10	0.09
3	Equivalent Bacterial Density	Approx. $1.5 \times 10^8$ CFU/mL	conforms
4	Stability Test	No precipitation or flocculation occurs during shelf life	pass

**LABOTIQ** Standard of Product Reference to NIST-traceable McFarland 0.5 reference standard and CLSI M07 & M02

SUDIRO S.Si.  
Manager QC