

TIQ-F-QC-001 Rev.00

Certificate Of Analysis

Product Name : ORP Standard Solution 1000 mV

Catalog Number : I-9087

Batch Number : 9087-00000000

CAS Number : Not available

Molecular Weight : Not available

Mol. Formula : Not available

Expire Date : 12 months



No	Test	Specifications	Results
1	Appearance	Clear yellow solution	Complies
2	ORP value at 25 °C	1000 ± 75 mV	998 mV
3	pH	acidic	Complies

LABOTIQ In-house product specification

ORP (Oxidation Reduction Potential) Standard Solution ,Prepared gravimetrically and measured using calibrated ORP meter with platinum electrode and Ag/AgCl reference electrode at 25°C.



SUDIRO S.Si.
Manager QC